

Pressure Test Certificate



UNIVERSITY
OF ABERDEEN



Oceanlab
Newburgh
Aberdeenshire
AB41 6AA

Contact:
David Sproule
Tel 01224 274403
d.sproule@abdn.ac.uk

Client:	Nortek Piezo Limited
Certificate Number	14-060
Date	12-05-2014
Purchase Order No	403

Serial Numbers	Equipment	Description of Test	Final Pressure (Bar)
007417, 007418, 007419, 007420, 007421, 007422, 007423, 007424, 007425.	9 off 55KHz Transducers.	<p>The items were placed in a hydrostatic pressure vessel and exposed to the following pressure cycle:</p> <ul style="list-style-type: none"> • Ramp up from 0 bar to 165 bar • Hold at pressure for 30 minutes. • Ramp down from 165 bar to 0 bar. <p>The items were removed from the vessel and visually inspected.</p> <p>Examination of the items revealed no apparent effect of hydrostatic testing.</p>	165
Signature of Test Operator			
Name (Block Capitals)		DAVID SPROULE	