



Statement of Compliance

Based on tests carried out in our Test Laboratory, at Høvik, Norway,

In the period 2015-09-17 to 2015-11-20,

Applica Test & Certification AS hereby state that the product

Signature 1000 Instrument

Manufactured by

Nortek AS

Vangkroken 2, 1351 Rud, Norway

complies with requirements given in EN 61000-6-2: 2005 and EN 61000-6-3: 2007/A1:2011/AC:2012.

The product has passed the following tests

Conducted Emission

Radiated Emission

Voltage Flicker

Harmonic Current

Conducted Radio Frequency Immunity

Radiated Electromagnetic Field Immunity

Fast Low Energy Transients / Bursts

Slow High Energy Transients / Surges

Voltage Dips & Interruptions

Electrostatic Discharge

Power-Frequency Magnetic Field

For more details, please refer to Applica Test & Certification Technical report no. 20715.

For Applica Test & Certification AS

Høvik, 2016-01-08

A handwritten signature in blue ink, reading 'Vibeke Vigmostad'.

Vibeke Vigmostad

Department Leader

A handwritten signature in blue ink, reading 'Michael Dinneen'.

Michael Dinneen

Project Manager

Notice: Any significant changes in design or construction may render this Statement of Compliance invalid.

HEAD OFFICE: Applica test and Certification AS, Veritasveien 1, 1363 Høvik, Norway. www.applica.no