



Statement of Compliance

Based on tests carried out in our Test Laboratory, at Høvik, Norway,
In the period 2016-05-09 to 2016-06-06,

Applica Test & Certification AS hereby state that the product

Signature 55 Instrument

Manufactured by

Nortek AS

Vangkroken 2, 1351 Rud, Norway

complies with requirements given in EN 61000-6-2: 2005 and EN 61000-6-3: 2007/A1:2011/AC:2012.

The product has passed the following tests

Conducted Emission

Radiated Emission

Voltage Flicker

Harmonic Current

Conducted Radio Frequency Immunity

Radiated Electromagnetic Field Immunity

Fast Low Energy Transients / Bursts

Slow High Energy Transients / Surges

Voltage Dips & Interruptions

Electrostatic Discharge

Power-Frequency Magnetic Field

For more details, please refer to Applica Test & Certification Technical report no. 20856.

For Applica Test & Certification AS
Høvik, 2016-07-05

A handwritten signature in blue ink that reads 'Vibeke Vigmostad'.

Vibeke Vigmostad
Department Leader

A handwritten signature in blue ink that reads 'Michael Dinneen'.

Michael Dinneen
Project Manager

Notice: Any significant changes in design or construction may render this Statement of Compliance invalid.

HEAD OFFICE: Applica test and Certification AS, Veritasveien 1, 1363 Høvik, Norway. www.applica.no